

https://www.geekcert.com/certified-hemodialysis-technician-cht.html 2024 Latest geekcert CERTIFIED-HEMODIALYSIS-TECHNICIAN-CHT PDF and VCE dumps Download

CERTIFIED-HEMODIALYSIS-TECHNICIAN-CHT^{Q&As}

BONENT Certified Hemodialysis Technician CHT

Pass BONENT CERTIFIED-HEMODIALYSIS-TECHNICIAN-CHT Exam with 100% Guarantee

Free Download Real Questions & Answers **PDF** and **VCE** file from:

https://www.geekcert.com/certified-hemodialysis-technician-cht.html

100% Passing Guarantee 100% Money Back Assurance

Following Questions and Answers are all new published by BONENT Official Exam Center https://www.geekcert.com/certified-hemodialysis-technician-cht.html 2024 Latest geekcert CERTIFIED-HEMODIALYSIS-TECHNICIAN-CHT PDF and VCE dumps Download



VCE & PDF

- Instant Download After Purchase
- 100% Money Back Guarantee
- 😳 365 Days Free Update
- 800,000+ Satisfied Customers





QUESTION 1

The major cause associated with the failure of the arteriovenous (AV) graft is

- A. aneurysm.
- B. rejection.
- C. thrombosis.
- D. steal syndrome.

Correct Answer: C

QUESTION 2

Which of the following is LEAST important in the orientation of a new employee?

- A. Developing a friendship with the orientee
- B. Supervising and evaluating the orientee\\'s clinical performance
- C. Adapting your teaching style to the orientee\\'s learning style
- D. Reviewing unit policies and procedures with the orientee

Correct Answer: A

QUESTION 3

Directly after treatment initiation, a patient complains of chest pain, shortness of breath, and burning the access arm. The patient is MOST likely experiencing

- A. a pyrogen reaction.
- B. a sterilant reaction.
- C. an air embolism.
- D. effects of blood flow that has been initiated too quickly.

Correct Answer: B

QUESTION 4

Which of the following may NOT be a common psychological reaction for patients beginning maintenance dialysis?

A. Guilt



- B. Anxiety
- C. Hostility
- D. Depression
- Correct Answer: C

QUESTION 5

The storage time of aqueous bicarbonate concentrate is limited because

- A. bacterial growth may occur.
- B. degassing of carbon dioxide solution.
- C. old bicarbonate will affect ulrafiltration (UF).
- D. bicarbonate growth causes leaching in the container.

Correct Answer: A

CERTIFIED-HEMODIALYSICERTIFIED-HEMODIALYSICERTIFIED-HEMODIALYSIS-TECHNICIAN-CHTS-TECHNICIAN-CHT StudyS-TECHNICIAN-CHT ExamPractice TestGuideQuestions